



WASHINGTON STATE PATROL  
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: April 16, 2006 INITIALS: DEC OWNER: 7/SP  
DISTRICT: 8 ADDRESS: \_\_\_\_\_  
LOCATION: 79th Shop COUNTY: \_\_\_\_\_  
MAKE: Hoana TYPE: port. wheel weigher ACCURACY CLASS: III  
MODEL: 101 ID. NO.: 24151 CAPACITY: 20,000 lbs SECTIONS: 1  
MINIMUM GRADUATIONS: 50 lb REG. ELEMENT: dial PLATFORM: 15" X 26"

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
Center	5000	5000	0				
"	10000	10000	0				
"	15000	15050	+50				
Center	20000	20000	0				
	25000						

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as is shown on the attached "Certificate of Accuracy", No. 19577

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Brown Date: 4/18/2006

State of Washington  
County of Thurston  
Signed and sworn to (or affirmed) before  
me on 4-18-06 by

(Signature of notary public)

Printed Name of Notary

My appointment expires



3000-362-021 (R) (10/98)

DISTRIBUTION: WHITE - COURT GREEN - DISTRICT CANARY - PROPERTY MANAGEMENT PINK & GOLDENROD - SERVICE TRUCKS

Service Technician: Mike Schuster

Form 1 03/01/00